

konformitätserklärung



17-10-09-463 rev.: -

Wir, der Hersteller, erklären in alleiniger Verantwortung, dass die unten aufgeführten Produkte den einschlägigen Bestimmungen der nachstehenden Verordnungen / Richtlinien entsprechen.

Basis UDI-DI	401203000000000000058R4
Device Model Basis UDI-DI	MDMST002elmf
Produkte	seca 286
REF	2861300004, 2861300204, 2861300904
Zweckbestimmung	Die elektronische Messstation unterstützt Ärzte bei Diagnose- oder Therapieentscheidungen, die auf dem Gewicht, der Größe oder Änderungen der gemessenen Parameter basieren.
Klassifizierung als Medizinprodukt	Klasse I mit Messfunktion
Konformitätsbewertungsverfahren (EU) 2017/745	Anhang IX
Bescheinigung (EU) 2017/745	G11 012163 0089

Verordnungen / Richtlinien:

(EU) 2017/745	Verordnung über Medizinprodukte
2011/65/EU	Richtlinie zur Beschränkung der Verwendung bestimmter gefährlicher Stoffe in Elektro- und Elektronikgeräten
2014/53/EU	Richtlinie über die Bereitstellung von Funkanlagen auf dem Markt

Hersteller:	seca gmbh & co. kg Hammer Steindamm 3-25 22089 Hamburg, Deutschland	Single Registration Number (SRN): DE-MF-000005469
--------------------	---	--

Made in China
Designed in Germany

Notifizierte Stelle / Benannte Stelle:	(EU) 2017/745: TÜV SÜD Product Service GmbH Ridlerstrasse 65 80339 München, Deutschland Kennnummer: 0123
---	--



Diese Konformitätserklärung ist gültig ab dem Datum der Unterzeichnung bis zur Ausstellung einer revidierten Konformitätserklärung aufgrund von Änderung der oben genannten Produkte.

Hamburg, 15. Juli 2025

Frederik Vogel
CEO Development & Manufacturing

Anhang

Angewandte Normen und Spezifikationen:

EN 60601-1	2006 + Cor.:2010 + A1:2013 + AC:2014 + A1:2013/AC:2014 + A12:2014 + A2:2021
EN 60601-1-2	2015 + A1:2021
EN ISO 62304	2006 + Cor.: 2008 + A1:2015
EN ISO 20417	2022
EN ISO 15223-1	2021
EN 300 328	V2.2.2

declaration of conformity



We, the manufacturer, declare in sole responsibility that the products mentioned below are in conformity with the respective regulations / directives of the following directives.

Basic UDI-DI	401203000000000000058R4
Device Model Basic UDI-DI	MDMST002elmf
Products	seca 286
REF	2861300004, 2861300204, 2861300904
Intended purpose	The electronic measuring station supports physicians in decision-making regarding diagnosis or therapy, based on the patient's weight, height or changes of the measured parameters.
Classification as a medical device	Class I with measuring function
Conformity assessment procedure (EU) 2017/745	Annex IX
Certificate (EU) 2017/745	G11 012163 0089

Regulations / Directives:

(EU) 2017/745	Regulation on medical devices
2011/65/EU	Directive on the restriction of the use of certain hazardous substances in electrical and electronic equipment
2014/53/EU	Directive relating to the making available on the market of radio equipment

Manufacturer: seca gmbh & co. kg
Hammer Steindamm 3-25
22089 Hamburg, Germany

Single Registration Number (SRN):
DE-MF-000005469

Made in China
Designed in Germany

**Notified body /
Notified bodies:** (EU) 2017/745:
TÜV SÜD Product Service GmbH
Ridlerstrasse 65
80339 München, Germany
Reference number: 0123



This declaration of conformity is valid from the date of signature until a revised declaration of conformity is issued due to modification of the above-mentioned products.

Hamburg, 15 July 2025

Frederik Vogel
CEO Development & Manufacturing

declaration of conformity



Annex

Applied standards and specifications:

EN 60601-1	2006 + Cor.:2010 + A1:2013 + AC:2014 + A1:2013/AC:2014 + A12:2014 + A2:2021
EN 60601-1-2	2015 + A1:2021
EN ISO 62304	2006 + Cor.: 2008 + A1:2015
EN ISO 20417	2022
EN ISO 15223-1	2021
EN 300 328	V2.2.2